

Sensor Certification

SENSIRION
THE SENSOR COMPANY

Product: Rapid RH[®] Product Family
Description: Digital Humidity and Temperature Sensor

This is to certify that above product is calibrated using dedicated equipment which is calibrated according to ISO/IEC 17025 and calibrated by comparison against natural standards, natural physical constants, consensus standards, or by ratio type measurements using self-calibrating techniques.

Standards incorporated to establish this comparison are NIST and NPL traceable.

NIST Traceable Accuracy Information: Range: 20%-90% RH, Accuracy: +/- 1.8% RH. Range: 90%-95% RH, Accuracy: +/- 3% RH

NIST
National Institute of
Standards and Technology



CEO, Sensirion AG

NPL
National Physical Laboratory



Quality Manager, Sensirion AG

Lot # 515000110

Calibration Date JUN 2015
515000110